



Gastro⁺™ Gastrolyzer® Maintenance Log

Calibration of the monitors should be performed every **3 months** regardless of use or when prompted by the monitor.

D-piece™ valves must be changed out **monthly**, for proper infection control.

Scheduled Date	Actual Date	Calibration	Cleaning	Change D-piece	Name Signature