

Smokerlyzer® Maintenance

Calibration of the monitors should be performed every <u>6 months</u> regardless of use or when prompted by the monitor.

D-piece[™] **valves** must be changed out **monthly**, for proper infection control.

Scheduled Date	Actual Date	Calibration	Cleaning	Change D-piece	Name Signature



www.covita.net | tel 800.707.5751 | fax 800.721.2377 212 Cottage Grove Ave. Suite C Santa Barbara, CA 93101