Smokerlyzer® Maintenance

Calibration of the monitors should be performed every **6 months** regardless of use or when prompted by the monitor.

D-piece valves must be changed out **monthly**, for proper infection control.

Scheduled Date	Actual Date	Calibration	Cleaning	Change D-piece	Name | Signature

© Copyright 2019 coVita LLC All Rights Reserved v091119| Reproduction permitted for organization use.